



2-Octanol Analysis

1. Analytical Condition

YL 6500 Series

Oven : 40°C(0.1min)->15°C/min->200°C(5min)

Column : HP-INNOWAX (30m*0.53mm*1.0um)

Carrier gas : N₂, 6 ml/min (Split ratio 10:1)

Injector : Capillary 250°C

Detector : FID 250°C

Injection Volume : 1ul (Liquid)

2. Chromatogram

