



HCl (Hydrogen chloride) Analysis

1. Analytical Condition

YL 6500 Series

Oven : 35°C(5min)->10°C/min->130°C(1min)

Column : SPB-5 (60m*0.32mm*1.0um)

Carrier gas : Ar, 3ml/min (Split ratio 10:1)

Injector : Capillary 150°C

Detector : TCD 150°C, Sen=4

Injection Volume : Gas 0.5ml

2. Chromatogram

