



Trace MeOH, 2-PrOH Analysis In EtOH

1. Analytical Condition

YL 6500GC Series

Oven : 40°C(8min)

Carrier gas : N₂, 3mL/min (Split ratio 100:1)

Injector : Capillary 250°C

Detector : FID 250°C

Column : HP-VOC(60 m*0.32 mm*1.8 μ m)

Injection Volume : 1 μ L

2. Chromatogram

